


DC Regulated Power Supplies

Model 382200 (1 Amp) and 382202 (3 Amp)



Introduction


Congratulations on your purchase of the Extech DC Regulated Power Supply. These units are solid state, compact, regulated supplies suitable for many applications including bench testing, field service, hobby and telecommunication equipment use. This meter is shipped fully tested and calibrated and, with proper use, will provide years of reliable service. . Please visit the Extech Instruments website (www.extech.com) to check for the latest version of this User Guide and to register this product.

Specifications

Model	382200	382202
Display	Dual LCD	
Display Accuracy (V)	$\pm(1.5\%rdg + 5D)$	
Display Accuracy (A)	$\pm(2.0\%rdg + 5D)$	
Voltage Output, DC	30.0 Volts	18.0 Volts
Current Output, DC	0 – 1.00 Amp	0 – 3.00 Amps
Ripple and Noise	<0.5mV/<3mA	
Line Regulation	< 0.01% + 3mV/< 0.2% + 3mA	
Load Regulation	< 0.01% + 2mV/< 0.2% + 3mA	
Power	115/230VAC (50/60Hz)	
Dimensions	9.4 x 4.25 x 6" (240 x 108 x 154mm) (W x H x D)	
Weight	4.4lbs. (2 kg)	6.6lbs. (3kg)

Meter Description

1. Voltage display
2. Current display
3. Negative (-) output terminal
4. Positive (+) output terminal
5. Illuminated On/Off switch
6. Current adjustment
7. Voltage adjustment



Operation

CAUTION: Verify that the AC voltage setting and fuse rating matches the power source being used BEFORE applying power to the instruments. Do not connect to a voltage source that is greater.

CAUTION: Limit full load operation (>25watts for the 382200 and >45watts for the 382202) to 15 minutes continuous operation. Always allow a cool down period after full load operation.

1. The Power Supply must be powered using the rated line voltage within $\pm 10\%$.
2. Before turning the power supply ON, remove any load and set the Voltage and Current Adjustment Knobs to their full counter-clockwise positions.
3. Use the Current and Voltage Adjustment knobs for setting variable Current and Voltage output signals respectively.
4. The LCD displays will indicate actual current and voltage outputs.
5. Keep the meter's cooling vents (top and sides) clear of obstacles to prevent overheating.

Current Limit

To avoid having the load draw too much current, a current limit can be set as follows:

1. With the voltage and current output adjustments set fully off, carefully short the positive and negative output terminals together.
2. Turn the Output Voltage knob clockwise to the 25% position (approx.).
3. Adjust the Output Current knob clockwise until the desired current level is displayed.
4. Turn the Output Voltage knob counter-clockwise (current display will go to 0.00).
5. Connect a load to the output terminals and adjust the voltage to the desired level.

Fuse Replacement

The fuse receptacle is located on the back of the unit next to the power cable input. Each unit comes with one 115V 2A fuse installed in the unit. Three spare fuses are included, a second 115V 2A fuse, and two 230V 1A fuses. See below for the fuse specifications.

	115V	230V
382200	2A	1A
382202	2A	1A

230V Operation

1. Replace the 2A (115V) fuse with a 1A (230V) fuse.
2. Toggle the switch located on the rear of the unit from 115V to 230V.
3. Use a power cord that matches the outlet for the country in question.

Warranty

FJRS Systems, Inc. warrants its Extech Instruments and devices to be free of defects in parts and workmanship for one year from date of shipment (a six month limited warranty applies to sensors and cables). It should become necessary to return the instrument for service during or beyond the warranty period, contact the Customer Service Department for authorization. Visit the website www.extech.com for contact information. A Return Authorization (RA) number must be issued before any product returned. The sender is responsible for shipping charges, freight, insurance and proper packaging to prevent damage in transit. This warranty does not apply to defects resulting from action of the user such as misuse, improper wiring, operation outside of specification, improper maintenance or repair, or unauthorized modification. FJRS Systems, Inc. specifically disclaims any implied warranties of merchantability or fitness for a specific purpose and will not be liable for any direct, indirect, incidental or consequential damages. FJRS total liability is limited to repair or replacement of the product. The warranty set forth above is inclusive and no other warranty, whether written or oral, is expressed or implied.

Calibration, Repair, and Customer Care Services

FJRS Systems, Inc. offers repair and calibration services for the Extech Instruments products we sell. NIST certification for most products is also provided. Call the Customer Service Department for information on calibration services available for this product. Annual calibrations should be performed to verify meter performance and accuracy. Technical support and general customer service is also provided, refer to the contact information provided below.

Support Lines: U.S. (877) 439-8324; International: +1 (603) 324-7800

Technical Support: Option 3; E-mail: support@extech.com

Repair & Returns: Option 4; E-mail: repair@extech.com

Product specifications are subject to change without notice

Please visit our website for the most up-to-date information

www.extech.com

FJRS Commercial Systems, Inc., 91 Townsend West, Nashua, NH 03063 USA

ISO 9001 Certified

Copyright © 2014-2015 FJRS Systems, Inc.

All rights reserved including the right of reproduction in whole or in part in any form

www.extech.com