



Precision Instruments Skyray Elaborates



ICP-2000

Sequential Inductively Coupled Plasma Spectrometer

Skyray Instrument Inc.
50 Bradstreet Hill Park, Suite 201,
Braintree, MA USA 02184
Tel: 617.202.3879 Fax: 781.539.4766
Website: www.skyrayinstrument.com



Precision Instruments Skyray Elaborates

Skyray Instrument Inc. is located in the scenery Braintree Hill Park, in Braintree, MA and is a high-technology enterprise specializing in the development, manufacturing and sales of analytical testing equipment in spectroscopy, chromatography and Mass Spectrometry.

Skyray has built a strong reputation and has gained over 80% of the market share in XRF technology within the past 5 years alone. Skyray Instrument has stepped up to a level of highly reputable equipment, with some of the largest manufacturing companies entrusting us with supplying them with our analytical equipment, Skyray has built a strong ever lasting reputation globally.

Skyray Instrument has recently joined the North American Market and wants to bring these low-cost, precise analytical equipment to the North American Consumer and Manufacturing market. The company possesses a first class expert management team, a strong Research Development Team, unparalleled technology with over 50 patents and a reputable service team. Skyray Instrument is the most trustworthy friend of the users as it implements rigid quality control on its products, pursues technological innovation ceaselessly and offers technical support promptly. Skyray Products command a good market both at home and abroad. Skyray representatives can be found in more than 50 countries world-wide including in regions such as Russia, Germany, Italy, Austria, Spain, Switzerland, Korea and Singapore and many more. There are over 100 offices and service stations all over the world.

Skyray Instrument, as an industry technology leader, is constantly exploring the pinnacle of analytical field and providing customers with more advanced products and more satisfied services, meanwhile, providing more perfect solutions for electronics, electric appliances, jewelry, toys, food, construction materials, metallurgy, minerals, plastics, petroleum, chemistry, medicine and other industries.



ICP-2000 Sequential Inductively Coupled Plasma Spectrometer



Application fields:

ICP-2000 is capable of measuring approximately 70 trace to constant elements in different materials. Wide application extends to geology, metallurgy, rare earth materials, electroplating, cement, petroleum, chemical engineering, environmental monitoring, non-ferrous metals, medical science and hygiene, food, agriculture, scientific research Institutes, colleges, and other industrial enterprises.

Main characteristics of ICP-AES:


- ◊ Simultaneous measurement of multi elements
- ◊ Low detection limit: 0.1-10ppb, and even lower
- ◊ Wide density linear range, 5-6 digits
- ◊ Least chemical interference, maximum accuracy, $RSD \leq 2\%$
- ◊ High analysis speed, low cost

www.skyrayinstrument.com

The Instrument Structure

The instrument consists of five parts:

- Radio Frequency Generator
- Sample Introduction System
- Optical System
- Photoelectric Conversion Device
- Computer



Radio Frequency Generator

RF generator's precision parts ensure stable output power; military electron tube offers increased reliability; improved feedback mechanism provides better stability.

RF Generator Specifications:

- 1). Circuit type: inductively feedback self-oscillation circuit, coaxial cable transmission, matched tuning, power feedback close loop auto control
- 2). Working frequency: 40MHz \pm 0.05%
- 3). Frequency stability: <0.1%
- 4). Power output: 800W 1200W
- 5). Power output stability: <0.3%
- 6). Intensity of electromagnetic leakage: electric field intensity: E: <2V/m @30cm form shell.



▲ High Frequency Box

High Frequency Box excites very stable plasma torch by exploiting the inductively feedback self-oscillation circuit. The circuit is transmitted by coaxial cable, matchingly tuned and implements close-loop auto-control by means of power feedback (National Patented Invention). Industrial argon can be used in this kind of generator, which not only makes the lighting the ICP torch easy but also saves the cost of highly purified argon.



▲ Match box

The high performance match box is composed of vacuum ceramic adjustable capacitors and gas adjustable capacitors, which allows it to reach the best match state and achieve maximum effective output on load coils.



▲ Directional coupler

The stability of power output is achieved by controlling the anode voltage of oscillation valve after the power signals are sent out of the directional coupler and get amplified by the amplifier.

Sample introduction system specifications:

- 1). Inner diameter of output coils: 25 mm, 3 coils
- 2). Torch: three concentric torches, outer diameter 20 mm, quartz
- 3). Outer diameter of coaxial nebulizer: 6mm
- 4). Outer diameter of double-pass spray chamber: 34mm
- 5). Specifications of Ar flowmeter and carrier gas pressure gauge:
 - ① Plasma gas flowmeter(100-1000) L/h (1.6-16L/min)
 - ② Auxiliary gas flowmeter (10-100) L/h (0.16-1.66L/min)
 - ③ Carrier gas flowmeter(10-100) L/h (0.16-1.66L/min)
 - ④ Carrier gas pressure stabilization valve (0-0.4MPa)
 - ⑤ Cooling water: Water temperature: 15-25°C Water pressure: >0.1MPa Resistivity>1M Ω

Sample introduction system

In this system, coaxial nebulizer, double-pass and cyclone spray chambers, high salt and conventional torch tubes, and their free and optimized combinations all contribute to the measurement of various kinds of samples. Meanwhile, the system is easy to clean, replace and install. A 2-channel 10-roller peristaltic pump uses one of its channels for sample introduction, resulting in uniform and stable sample introduction, and the other channel for pumping away the waste, minimizing the influence the waste peristaltic has had on the sample introduction. Direct-drive and digital, the peristaltic pump delivers ease and simplicity of operation.


Both R.F generator and plasma load coils are cooled by circulated cooling water. ICP-2000 is equipped with the in-house designed AC-2500 auto temperature control water cooling system, which is characterized by good cooling results, strong cooling capability, wide range temperature control and low noise.



▲ Auto temperature control water cooling system

Optical system

The optical system utilizes the horizontal asymmetric optical path to reduce the width of the system and eliminate the interference of the stray lights. With adoption of large-area grating with long focal length, the relative aperture is broadened and sensitivity enhanced. At the same time, by using special metal material and finely-lapped fixed slits, it delivers superior measurement results in great precision.



▲ Grating drive



▲ Optical system

Optical system specifications:

- 1). Optical path: Czerny-Turner
- 2). Focal length: 1000 mm
- 3). Grating specification: Holographic ion beam etched grating, Number of grooves: 3600 grooves/mm
Area of grooves: (80 \times 110) mm
(Optional grating with number of grooves: 2400 grooves/mm and area of grooves: 80 \times 110)
- 4). Reciprocal linear dispersion: 0.26nm/mm
- 5). Resolution: \leq 0.008nm
- 6). Wavelength range:
 - 3600 grooves/mm for 195 to 500nm
 - 2400 grooves/mm for 195 to 800nm
- 7). Minimum step wavelength: 0.0006 nm
- 8). Exit, entrance slits: 20 μ m
- 9). Reflector: (78 \times 105 \times 16)mm
- 10). Transmission mirror: ϕ 30, 1:1 imaging
- 11). Temperature control: 32 \pm 1 $^{\circ}$ C

By using the imported precise and reliable step motor as the drive device of the grating, the system offers stable mechanical transmission, resulting in good stability and repeatability. The rotation of the grating is driven by the fine-pitch worm gear directly, ensuring high precision and minimal accumulated error. The optical system offers the ultimate in temperature 32 \pm 1 $^{\circ}$ C with the thermostatic device.


www.skyrayinstrument.com

Photoelectric Conversion Device:

- 1). Photomultiplier specifications: R212UH or R928
- 2). Negative HV power supply of photomultiplier: 200-1000V
Stability: <0.05%
- 3). Measurement range of photomultiplier current: 1nA-100 μ A
- 4). Signal collection is in V/F transfer: 1mV corresponds to 100Hz

Photoelectric Conversion Device

Precise control circuit and electronic measuring system includes Communication Interfaces System, Photomultiplier HV Power Supply and Control, Step Motor Control, and Signal Collection. The photomultiplier and components in use are new and high sensitive with wide strength range. The signal sampling is V/F converting, which has wide range of the sample content, high sensitivity and exceptional stability.






▲ Electronic control board

Personal computer

Personal computer configurations:

- 1). Brand name: FOUNDER Memory :512MB Hard disc: 80GB
- 2). Display: 17-inch TFT/LCD
- 3). Printer: Panasonic KX-P1121 24-pin printer

Functions of software




Zero-order searching is the foundation of preparing the instrument for measurement positioning. You are suggested to perform a calibration using a blank to determine the mechanical zero point every time the instrument is turned on.

To develop methods, select the element of interest from the Periodic Table of Elements on the method interface according to its type after the program is started and select the line of the element.

The spectrum analysis mode performs auto peak searching of the added spectrum with the help of the software. When several peaks coexist, the users can manually identify the peak so as to eliminate the interference produced by elements other than the tested one.

Test example




Spectrum of As in waste water



Industrial waste water

Element	Wavelength	Sample content	Relative deviation
As	193.656	5.03 ppm	0.54%
As	228.812	5.03 ppm	0.44%
As	234.984	5.05 ppm	0.10%
As	278.022	5.00 ppm	0.31%


Test results of As



Prepare work curve

Element	Wavelength	Sample content	Relative deviation
As	193.656	5.03 ppm	0.54%
As	228.812	5.03 ppm	0.44%
As	234.984	5.05 ppm	0.10%
As	278.022	5.00 ppm	0.31%

For sample measurement, the concentration units are in % or ppm or µg/ml.



Data and results can be printed out or saved in copies.

www.skyrayinstrument.com

ICP-2000 Measurable elements and detection limits

Below 10PPb
 10-100PPb
 100-1000PPb
 Above 1000PPb
 Inmeasurable elements

H	2a	3a	4a	5a	6a	7a	1a	2b	3b	4b	5a	6a	7a	0				
He																		
Li	Be	<i>Skyray Skyray Instrument</i>										B	C	N	O	F	Ne	
Na	Mg											Al	Si	P	S	Cl	Ar	
K	Ca	Sc	Ti	V	Cr	Mn	Fe	Co	Ni	Cu	Zn	Ga	Ge	As	Se	Br	Kr	
Rb	Sr	Y	Zr	Nb	Mo	Tc	Ru	Rh	Pd	Ag	Cd	In	Sn	Sb	Te	I	Xe	
Cs	Ba	La	Hf	Ta	W	Re	Os	Ir	Pt	Au	Hg	Tl	Pb	Bi	Po	At	Rn	
Fr	Ra	Ac																
		La	Ce	Pr	Nd	Pm	Sm	Eu	Gd	Tb	Dy	Ho	Er	Tm	Yb	Lu		
		Ac	Th	Pa	U	Np	Pu	Am	Cm	Bk	Cf	Es	Fm	Md	No	Lw		